

## T 32 X-ray Tank Developer

### Description :

- developer for (manual) tank processing of X - ray films;
- to process both medical and technical X - ray films with high contrast at good utilization of sensitivity;
- use as tank solution and replenisher;
- to prepare a fresh tank solution no starter is needed;
- high durability;
- liquid concentrate;
- one - part.

### Processing Conditions :




developing temperature :	18 °C	developing time :	5 - 6 min.
	20 °C		4 - 5 min.
	22 °C		3 - 4 min.
	24 °C		2 - 3 min.

### Replenishment :

- For processing of 1 m<sup>2</sup> X - ray film 300 to 400 ml tank solution are necessary at standard conditions .




### Mix Instruction :

preparation of tank developer or replenisher:

Water ( 25 - 30 °C )	T 32 - Concentrate	= tank solution
 about 3/4 of tank volume		
4,0 L	1 L	5,0 L
20,0 L	5 L	25,0 L

**T 32 X-ray Tank Developer**

Preparation of replenisher solution:

Water ( 15 - 30 °C )	T 32 - ConCentrate	= Replenisher solution
	 (add under permanent stirring)	
2,5 L	1 L	3,5 L
17,5 L	5 L	17,5 L

**Assortment :**

Order Number	Product Name	Mix Size to make ...	Capacity ( m <sup>2</sup> )
10040	T 32	10 x 5,0 Litres (cont. 10 x 1 L conc.)	90 - 120
10042	T 32	2 x 25,0 Litres (cont. 2 x 5 L conc.)	90 - 120

**Storage and Handling :**

- Until the preparation T 32 has to be kept in the closed original package and stored at temperatures of 10 to 25 ° C ..
- All users of photographic processing solutions should exercise the greatest care to avoid the chemicals contacting the skin or eyes . Always wear solution resistant gloves .
- The disposal of used photographic processing solutions has to be occurred in accordance with the legal regulations .